

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U101219\
 Data File : VU035119.D
 Acq On : 12 Oct 2019 04:05
 Operator : JC/SP
 Sample : K5309-19
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB-191007-01

Quant Time: Oct 14 01:43:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Oct 14 01:16:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.87	114	171776	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	179387	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	67304	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	51380	5.06	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.20%
7) Chloroethane-d5	1.68	69	48349	5.10	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.00%
11) 1,1-Dichloroethene-d2	2.26	63	74522	3.48	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.60%
20) 2-Butanone-d5	4.17	46	151057	50.52	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.04%
24) Chloroform-d	4.63	84	92174	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.20%
26) 1,2-Dichloroethane-d4	5.29	65	54905	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.60%
32) Benzene-d6	5.32	84	192712	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.60%
36) 1,2-Dichloropropane-d6	6.31	67	66353	4.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.40%
41) Toluene-d8	7.55	98	167806	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.40%
43) trans-1,3-Dichloropropene-	7.84	79	19181	3.79	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.80%
46) 2-Hexanone-d5	8.30	63	102084	42.41	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.82%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	56358	4.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.20%
64) 1,2-Dichlorobenzene-d4	11.85	152	56992	5.02	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.40%

Target Compounds

					Ovalue
3) Chloromethane	1.32	50	6834	0.397 ug/L	97
13) Acetone	2.31	43	17193	7.020 ug/L	100
25) Chloroform	4.65	83	9357	0.355 ug/L	90
42) Toluene	7.62	91	24700	0.438 ug/L	93
53) m,p-Xylene	9.36	106	3850	0.167 ug/L	83
55) Styrene	9.78	104	6363	0.175 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101219\
Data File : VU035119.D
Acq On : 12 Oct 2019 04:05
Operator : JC/SP
Sample : K5309-19
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB-191007-01

Quant Time: Oct 14 01:43:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Oct 14 01:16:14 2019
Response via : Initial Calibration

