

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101219\
 Data File : VU035122.D
 Acq On : 12 Oct 2019 05:17
 Operator : JC/SP
 Sample : K5342-01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AY7

Quant Time: Oct 14 01:56:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Oct 14 01:16:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	169783	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	174620	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	68390	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	51380	5.12	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.40%
7) Chloroethane-d5	1.67	69	49233	5.25	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.00%
11) 1,1-Dichloroethene-d2	2.26	63	74758	3.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.60%
20) 2-Butanone-d5	4.18	46	152963	51.75	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.50%
24) Chloroform-d	4.62	84	91115	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.20%
26) 1,2-Dichloroethane-d4	5.29	65	56710	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.00%
32) Benzene-d6	5.32	84	190879	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.20%
36) 1,2-Dichloropropane-d6	6.31	67	69046	5.04	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.80%
41) Toluene-d8	7.55	98	162638	4.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.20%
43) trans-1,3-Dichloropropene-	7.84	79	20152	4.09	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.80%
46) 2-Hexanone-d5	8.30	63	107126	45.72	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.44%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	57484	4.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.60%
64) 1,2-Dichlorobenzene-d4	11.84	152	59121	5.12	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.40%

Target Compounds					Ovalue
13) Acetone	2.31	43	3014	1.245 ug/L	99
47) Tetrachloroethene	8.21	164	5281	0.481 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101219\
Data File : VU035122.D
Acq On : 12 Oct 2019 05:17
Operator : JC/SP
Sample : K5342-01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AY7

Quant Time: Oct 14 01:56:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Oct 14 01:16:14 2019
Response via : Initial Calibration

