

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U101220\
 Data File : VU040654.D
 Acq On : 12 Oct 2020 13:05
 Operator : SY/MD
 Sample : L4345-04
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YAQX6

Quant Time: Oct 13 05:07:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 13 04:32:46 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	127288	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	131518	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	55442	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	43232	4.10	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.00%
7) Chloroethane-d5	1.92	69	38735	4.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.20%
11) 1,1-Dichloroethene-d2	2.58	63	68391	3.32	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.40%
20) 2-Butanone-d5	4.66	46	175632	52.04	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.08%
24) Chloroform-d	5.08	84	89592	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.80%
26) 1,2-Dichloroethane-d4	5.72	65	53672	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.20%
32) Benzene-d6	5.74	84	163147	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.00%
36) 1,2-Dichloropropane-d6	6.71	67	55759	4.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.80%
41) Toluene-d8	7.91	98	131204	4.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.60%
43) trans-1,3-Dichloropropene-	8.19	79	22484	4.51	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.20%
46) 2-Hexanone-d5	8.64	63	121491	47.26	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.52%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	47758	4.52	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	52430	5.10	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.00%

Target Compounds					Ovalue
13) Acetone	2.67	43	82090	31.377 ug/L	91
25) Chloroform	5.11	83	13202	0.658 ug/L	90
34) Trichloroethene	6.56	95	28812	2.631 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101220\
Data File : VU040654.D
Acq On : 12 Oct 2020 13:05
Operator : SY/MD
Sample : L4345-04
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YAQX6

Quant Time: Oct 13 05:07:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 13 04:32:46 2020
Response via : Initial Calibration

