

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101223\
 Data File : VU055760.D
 Acq On : 12 Oct 2023 12:38
 Operator : MD/SY
 Sample : 04792-04DL 100X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EY1B2DL

Quant Time: Oct 12 22:18:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM101023WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 12 22:16:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	448413	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	454681	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	212078	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	152437	37.900	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	75.800%		
7) Chloroethane-d5	1.885	69	129591	39.416	ug/L	-0.02
Spiked Amount 50.000	Range 70 - 130		Recovery =	78.840%		
11) 1,1-Dichloroethene-d2	2.557	63	216142	30.139	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	60.280%		
21) 2-Butanone-d5	4.612	46	237240	79.603	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	79.600%		
24) Chloroform-d	5.055	84	332701	43.148	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	86.300%		
26) 1,2-Dichloroethane-d4	5.695	65	223815	47.158	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.320%		
32) Benzene-d6	5.721	84	642616	44.173	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	88.340%		
36) 1,2-Dichloropropane-d6	6.685	67	206798	44.902	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	89.800%		
41) Toluene-d8	7.894	98	550916	41.939	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	83.880%		
43) trans-1,3-Dichloroprop...	8.174	79	95188	43.026	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	86.060%		
47) 2-Hexanone-d5	8.631	63	144146	79.615	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	79.610%		
56) 1,1,2,2-Tetrachloroeth...	10.753	84	335053	47.264	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	94.520%		
66) 1,2-Dichlorobenzene-d4	12.190	152	213107	50.636	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	101.280%		
Target Compounds						Qvalue
20) cis-1,2-Dichloroethene	4.660	96	13525	3.646	ug/L	97
34) Trichloroethene	6.544	95	3148	0.800	ug/L #	64
46) Tetrachloroethene	8.550	164	151459	53.792	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101223\
Data File : VU055760.D
Acq On : 12 Oct 2023 12:38
Operator : MD/SY
Sample : 04792-04DL 100X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EY1B2DL

Quant Time: Oct 12 22:18:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM101023WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 12 22:16:59 2023
Response via : Initial Calibration

