

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101320\
 Data File : VU040674.D
 Acq On : 13 Oct 2020 11:42
 Operator : SY/MD
 Sample : VU1013WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK27

Quant Time: Oct 14 06:12:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 14 05:57:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	116333	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	119142	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	49650	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	43053	4.47	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.40%
7) Chloroethane-d5	1.92	69	38673	5.20	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.00%
11) 1,1-Dichloroethene-d2	2.58	63	68631	3.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.80%
20) 2-Butanone-d5	4.66	46	177154	57.43	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.86%
24) Chloroform-d	5.08	84	90444	5.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.00%
26) 1,2-Dichloroethane-d4	5.72	65	55442	5.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.60%
32) Benzene-d6	5.74	84	161800	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.80%
36) 1,2-Dichloropropane-d6	6.70	67	56446	5.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.80%
41) Toluene-d8	7.91	98	127212	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.40%
43) trans-1,3-Dichloropropene-	8.19	79	21168	4.68	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.60%
46) 2-Hexanone-d5	8.64	63	118975	51.09	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.18%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	49558	5.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	50254	5.46	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.20%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	2268	0.213 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101320\
Data File : VU040674.D
Acq On : 13 Oct 2020 11:42
Operator : SY/MD
Sample : VU1013WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK27

Quant Time: Oct 14 06:12:24 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 14 05:57:05 2020
Response via : Initial Calibration

