

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101321\
 Data File : VU045306.D
 Acq On : 13 Oct 2021 13:57
 Operator : SY/MD
 Sample : VU1013WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1013WBL01

Quant Time: Oct 14 06:16:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U101321DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Oct 14 05:46:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.121	96	39265	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.639	95	13181	0.914	ug/l	0.00
Spiked Amount 1.000			Recovery	=	91.000%	
68) 1,2-Dichlorobenzene-d4	12.201	152	12361	0.916	ug/l	0.00
Spiked Amount 1.000			Recovery	=	92.000%	
Target Compounds						
15) Methylene Chloride	3.051	84	4898	0.396	ug/l	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101321\
Data File : VU045306.D
Acq On : 13 Oct 2021 13:57
Operator : SY/MD
Sample : VU1013WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1013WBL01

Quant Time: Oct 14 06:16:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U101321DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Oct 14 05:46:27 2021
Response via : Initial Calibration

