

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101322\
 Data File : VU051409.D
 Acq On : 13 Oct 2022 14:43
 Operator : JC/MD
 Sample : N5026-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Oct 14 01:22:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 14 01:20:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	188805	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	206723	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	70326	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	84760	5.565	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	111.400%
7) Chloroethane-d5	1.909	69	67022	5.122	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.400%
11) 1,1-Dichloroethene-d2	2.562	65	31236	5.188	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	103.800%
20) 2-Butanone-d5	4.642	46	170916	50.972	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.940%
24) Chloroform-d	5.060	84	137384	4.839	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
26) 1,2-Dichloroethane-d4	5.703	65	74944	5.118	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.400%
32) Benzene-d6	5.726	84	240156	4.223	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.400%
36) 1,2-Dichloropropane-d6	6.690	67	87733	4.542	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.800%
41) Toluene-d8	7.896	98	170986	3.361	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	67.200%#
43) trans-1,3-Dichloroprop...	8.179	79	26913	4.241	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.800%
46) 2-Hexanone-d5	8.636	63	95719	44.758	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.520%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	72434	4.695	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.800%
66) 1,2-Dichlorobenzene-d4	12.192	152	54889	4.490	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.800%
Target Compounds						Qvalue
16) Methylene chloride	3.044	84	17424	0.857	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101322\
Data File : VU051409.D
Acq On : 13 Oct 2022 14:43
Operator : JC/MD
Sample : N5026-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Oct 14 01:22:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 14 01:20:41 2022
Response via : Initial Calibration

