

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101322\
 Data File : VU051417.D
 Acq On : 13 Oct 2022 19:08
 Operator : JC/MD
 Sample : N5024-12
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Oct 14 01:24:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 14 01:20:41 2022
 Response via : Initial Calibration

Compound			R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards								
1) 1,4-Difluorobenzene			6.247	114	190357	5.000	ug/L	0.00
28) Chlorobenzene-d5			9.417	117	211281	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4			11.812	152	72070	5.000	ug/L	0.00
System Monitoring Compounds								
4) Vinyl Chloride-d3			1.594	65	86813	5.654	ug/L	0.00
Spiked Amount 5.000			Range 40 - 130		Recovery	= 113.000%		
7) Chloroethane-d5			1.909	69	72118	5.466	ug/L	0.00
Spiked Amount 5.000			Range 65 - 130		Recovery	= 109.400%		
11) 1,1-Dichloroethene-d2			2.562	65	29004	4.778	ug/L	0.00
Spiked Amount 5.000			Range 60 - 125		Recovery	= 95.600%		
20) 2-Butanone-d5			4.639	46	185011	54.726	ug/L	0.00
Spiked Amount 50.000			Range 40 - 130		Recovery	= 109.460%		
24) Chloroform-d			5.063	84	145211	5.073	ug/L	0.00
Spiked Amount 5.000			Range 70 - 125		Recovery	= 101.400%		
26) 1,2-Dichloroethane-d4			5.700	65	79898	5.412	ug/L	0.00
Spiked Amount 5.000			Range 70 - 130		Recovery	= 108.200%		
32) Benzene-d6			5.726	84	255965	4.404	ug/L	0.00
Spiked Amount 5.000			Range 70 - 125		Recovery	= 88.000%		
36) 1,2-Dichloropropane-d6			6.690	67	94763	4.801	ug/L	0.00
Spiked Amount 5.000			Range 60 - 140		Recovery	= 96.000%		
41) Toluene-d8			7.896	98	184570	3.549	ug/L	0.00
Spiked Amount 5.000			Range 70 - 130		Recovery	= 71.000%		
43) trans-1,3-Dichloroprop...			8.179	79	27723	4.275	ug/L	0.00
Spiked Amount 5.000			Range 55 - 130		Recovery	= 85.400%		
46) 2-Hexanone-d5			8.636	63	106843	48.882	ug/L	0.00
Spiked Amount 50.000			Range 45 - 130		Recovery	= 97.760%		
56) 1,1,2,2-Tetrachloroeth...			10.755	84	79169	5.021	ug/L	0.00
Spiked Amount 5.000			Range 65 - 120		Recovery	= 100.400%		
66) 1,2-Dichlorobenzene-d4			12.192	152	61756	4.930	ug/L	0.00
Spiked Amount 5.000			Range 80 - 120		Recovery	= 98.600%		
Target Compounds								Qvalue
16) Methylene chloride			3.044	84	18669	0.911	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101322\
Data File : VU051417.D
Acq On : 13 Oct 2022 19:08
Operator : JC/MD
Sample : N5024-12
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Oct 14 01:24:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 14 01:20:41 2022
Response via : Initial Calibration

