

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101422\
 Data File : VU051425.D
 Acq On : 14 Oct 2022 17:09
 Operator : JC/MD
 Sample : N4817-01DL 4X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 38025-110221DL

Quant Time: Oct 15 06:10:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U100422DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Wed Oct 05 02:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.115	96	46658	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.635	95	11608	0.757	ug/l	0.00
Spiked Amount 1.000			Recovery	=	76.000%	
68) 1,2-Dichlorobenzene-d4	12.192	152	12761	0.749	ug/l	0.00
Spiked Amount 1.000			Recovery	=	75.000%	
Target Compounds						
						Qvalue
15) Methylene Chloride	3.041	84	34995	1.432	ug/l	97
27) Chloroform	5.086	83	94237	3.503	ug/l	98
44) Bromodichloromethane	7.105	83	208875	10.823	ug/l	97
55) Dibromochloromethane	8.809	129	61623	4.715	ug/l	100
67) Bromoform	10.288	173	20700	2.744	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101422\
Data File : VU051425.D
Acq On : 14 Oct 2022 17:09
Operator : JC/MD
Sample : N4817-01DL 4X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
38025-110221DL

Quant Time: Oct 15 06:10:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U100422DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Wed Oct 05 02:45:21 2022
Response via : Initial Calibration

