

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101424\
 Data File : VU061128.D
 Acq On : 14 Oct 2024 10:46
 Operator : MD/SY
 Sample : P4380-20
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Oct 15 05:37:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 15 05:36:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	163718	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	147303	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	66238	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	45877	3.919	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.400%
7) Chloroethane-d5	1.907	69	40758	4.652	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.000%
11) 1,1-Dichloroethene-d2	2.560	65	21595	4.182	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.600%
20) 2-Butanone-d5	4.624	46	146511	52.255	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.500%
24) Chloroform-d	5.052	84	101477	4.919	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.400%
26) 1,2-Dichloroethane-d4	5.692	65	54955	5.026	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.600%
32) Benzene-d6	5.718	84	197613	5.088	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.800%
36) 1,2-Dichloropropane-d6	6.682	67	69088	5.416	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.400%
41) Toluene-d8	7.891	98	175562	4.836	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.800%
43) trans-1,3-Dichloroprop...	8.174	79	25405	4.658	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.200%
46) 2-Hexanone-d5	8.627	63	123311	54.625	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.260%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	50226	5.369	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	57773	5.455	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101424\
Data File : VU061128.D
Acq On : 14 Oct 2024 10:46
Operator : MD/SY
Sample : P4380-20
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Oct 15 05:37:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 15 05:36:38 2024
Response via : Initial Calibration

