

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101424\
 Data File : VU061129.D
 Acq On : 14 Oct 2024 11:10
 Operator : MD/SY
 Sample : P4389-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Oct 15 05:38:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 15 05:36:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	156740	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	142273	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	63089	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	45738	4.082	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.600%
7) Chloroethane-d5	1.907	69	40967	4.884	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.600%
11) 1,1-Dichloroethene-d2	2.560	65	21342	4.317	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.400%
20) 2-Butanone-d5	4.621	46	151414	56.408	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.820%
24) Chloroform-d	5.052	84	102304	5.180	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.600%
26) 1,2-Dichloroethane-d4	5.695	65	54866	5.241	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.800%
32) Benzene-d6	5.717	84	194462	5.184	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.600%
36) 1,2-Dichloropropane-d6	6.682	67	67246	5.458	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.200%
41) Toluene-d8	7.891	98	170105	4.851	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.000%
43) trans-1,3-Dichloroprop...	8.174	79	25183	4.781	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.600%
46) 2-Hexanone-d5	8.627	63	125700	57.652	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	115.300%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	50429	5.582	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	57188	5.669	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101424\
Data File : VU061129.D
Acq On : 14 Oct 2024 11:10
Operator : MD/SY
Sample : P4389-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Oct 15 05:38:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 15 05:36:38 2024
Response via : Initial Calibration

