

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U101518\
 Data File : VU027661.D
 Acq On : 15 Oct 2018 12:51
 Operator : MD/SY
 Sample : J5403-18
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Oct 16 07:52:22 2018
 Quant Title :
 QLast Update : Tue Oct 16 07:44:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	190173	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	174639	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	69227	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	61298	40.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.92%
7) Chloroethane-d5	1.67	69	53712	46.59	ug/L	-0.01
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.18%
11) 1,1-Dichloroethene-d2	2.27	63	91722	32.67	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.34%
21) 2-Butanone-d5	4.18	46	128706	106.53	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.53%
24) Chloroform-d	4.65	84	116736	48.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.54%
26) 1,2-Dichloroethane-d4	5.31	65	85593	51.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.06%
32) Benzene-d6	5.34	84	236871	47.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.66%
36) 1,2-Dichloropropane-d6	6.33	67	87848	51.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.48%
41) Toluene-d8	7.57	98	191971	42.91	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.82%
43) trans-1,3-Dichloropropene-	7.85	79	35786	43.32	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.64%
47) 2-Hexanone-d5	8.31	63	85794	99.08	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.08%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	114042	50.87	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.74%
64) 1,2-Dichlorobenzene-d4	11.86	152	69097	51.02	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.04%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101518\
Data File : VU027661.D
Acq On : 15 Oct 2018 12:51
Operator : MD/SY
Sample : J5403-18
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Oct 16 07:52:22 2018
Quant Title :
QLast Update : Tue Oct 16 07:44:54 2018
Response via : Initial Calibration

