

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U101520\
 Data File : VU040717.D
 Acq On : 15 Oct 2020 13:59
 Operator : SY/MD
 Sample : L4401-12
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Oct 16 05:01:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 16 04:23:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	110526	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	115723	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	49785	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	43481	4.75	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.00%
7) Chloroethane-d5	1.92	69	38865	5.50	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.00%
11) 1,1-Dichloroethene-d2	2.58	63	68704	3.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.80%
20) 2-Butanone-d5	4.65	46	173392	59.16	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.32%
24) Chloroform-d	5.08	84	89968	5.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.00%
26) 1,2-Dichloroethane-d4	5.72	65	56390	5.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.20%
32) Benzene-d6	5.74	84	166379	5.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.60%
36) 1,2-Dichloropropane-d6	6.70	67	57156	5.52	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.40%
41) Toluene-d8	7.91	98	132427	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.80%
43) trans-1,3-Dichloropropene-	8.19	79	21302	4.85	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.00%
46) 2-Hexanone-d5	8.64	63	119756	52.94	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.88%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	50161	5.40	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	54014	5.85	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.00%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	1551	0.153 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101520\
Data File : VU040717.D
Acq On : 15 Oct 2020 13:59
Operator : SY/MD
Sample : L4401-12
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Oct 16 05:01:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 16 04:23:22 2020
Response via : Initial Calibration

