

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U101520\
 Data File : VU040728.D
 Acq On : 15 Oct 2020 18:14
 Operator : SY/MD
 Sample : L4421-08
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFSJ8

Quant Time: Oct 16 05:39:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 16 04:23:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	108427	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	116466	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	47690	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	41235	4.59	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.80%
7) Chloroethane-d5	1.92	69	37526	5.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.40%
11) 1,1-Dichloroethene-d2	2.58	63	65392	3.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.40%
20) 2-Butanone-d5	4.66	46	167800	58.36	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.72%
24) Chloroform-d	5.08	84	86119	5.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.20%
26) 1,2-Dichloroethane-d4	5.72	65	52289	5.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.80%
32) Benzene-d6	5.74	84	157657	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.40%
36) 1,2-Dichloropropane-d6	6.70	67	53995	5.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.60%
41) Toluene-d8	7.91	98	121252	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.20%
43) trans-1,3-Dichloropropene-	8.19	79	20847	4.72	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.40%
46) 2-Hexanone-d5	8.64	63	114384	50.24	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.48%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	49047	5.24	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	51679	5.84	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.80%

Target Compounds					Ovalue
3) Chloromethane	1.53	50	5525	0.601 ug/L	99
13) Acetone	2.66	43	9653	4.331 ug/L	65
47) Tetrachloroethene	8.56	164	14491	2.094 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101520\
Data File : VU040728.D
Acq On : 15 Oct 2020 18:14
Operator : SY/MD
Sample : L4421-08
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFSJ8

Quant Time: Oct 16 05:39:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 16 04:23:22 2020
Response via : Initial Calibration

