

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101521\
 Data File : VU045316.D
 Acq On : 15 Oct 2021 11:35
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Oct 16 01:09:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U101321DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Oct 14 05:46:27 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 i	Fluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.382	0.387	-1.3	101	0.00
3 t	Chloromethane	0.398	0.402	-1.0	104	0.00
4 Rt	Vinyl Chloride	0.413	0.420	-1.7	101	0.00
5 T	Bromomethane	0.227	0.212	6.6	96	0.00
6 T	Chloroethane	0.268	0.253	5.6	99	0.00
7 T	Trichlorofluoromethane	0.481	0.469	2.5	98	0.00
8	1,1,2-Trichloro-1,2,2-trifl	0.278	0.290	-4.3	103	0.00
9 Rt	1,1-Dichloroethene	0.260	0.264	-1.5	101	0.00
10 t	Iodomethane	0.272	0.297	-9.2	96	0.00
11 t	Allyl Chloride	0.382	0.407	-6.5	103	0.00
12 t	Acrylonitrile	0.066	0.064	3.0	92	0.00
13 T	Acetone	0.079	0.069	12.7	89	0.00
14 T	Carbon Disulfide	0.793	0.797	-0.5	99	0.00
15 RT	Methylene Chloride	0.315	0.304	3.5	101	0.00
16 RT	trans-1,2-Dichloroethene	0.279	0.287	-2.9	101	0.00
17 t	1,1-Dichloroethane	0.524	0.552	-5.3	104	0.00
18 T	2-Butanone	0.101	0.111	-9.9	106	0.00
19	Cyclohexane	0.485	0.503	-3.7	101	0.00
20	Methylcyclohexane	0.482	0.502	-4.1	99	0.00
21 T	2,2-Dichloropropane	0.446	0.464	-4.0	103	0.00
22 RT	cis-1,2-Dichloroethene	0.307	0.321	-4.6	102	0.00
23 t	Diethyl Ether	0.246	0.241	2.0	98	0.00
24 t	tert-Butyl Alcohol	0.019	0.016	15.8	92	0.00
25 t	Methyl tert-Butyl Ether	0.726	0.710	2.2	96	0.00
26 t	Bromochloromethane	0.131	0.130	0.8	99	0.00
27 t	Chloroform	0.537	0.545	-1.5	103	0.00
28 RT	1,1,1-Trichloroethane	0.440	0.455	-3.4	102	0.00
29 T	1,1-Dichloropropene	0.402	0.414	-3.0	101	0.00
30 RT	Carbon Tetrachloride	0.341	0.353	-3.5	101	0.00
31 t	Isopropyl Ether	0.830	0.882	-6.3	103	0.00
32	Ethyl-t-butyl ether	0.836	0.850	-1.7	98	0.00
33	Tert-Amyl methyl ether	0.740	0.757	-2.3	98	0.00
34 t	Propionitrile	0.025	0.021	16.0	82	0.00
35 RT	Benzene	1.162	1.209	-4.0	103	0.00
36 RT	1,2-Dichloroethane	0.371	0.367	1.1	99	0.00
37 RT	Trichloroethene	0.283	0.289	-2.1	99	0.00
38 Rt	1,2-Dichloropropane	0.301	0.316	-5.0	102	0.00
39 t	Methacrylonitrile	0.099	0.094	5.1	94	0.00
40 t	Methyl acrylate	0.165	0.171	-3.6	94	0.00
41 t	Tetrahydrofuran	0.057	0.055	3.5	99	0.00
42 t	1-Chlorobutane	0.577	0.600	-4.0	103	0.00
43 T	Dibromomethane	0.150	0.151	-0.7	99	0.00
44 T	Bromodichloromethane	0.373	0.393	-5.4	101	0.00
45 T	4-Methyl-2-Pentanone	0.204	0.202	1.0	96	0.00
46 t	t-1,4-Dichloro-2-butene	0.072	0.081	-12.5	99	0.00
47 t	Methyl methacrylate	0.152	0.152	0.0	95	0.00
48 t	Ethyl methacrylate	0.300	0.306	-2.0	94	0.00
49 Rt	Toluene	0.715	0.742	-3.8	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101521\
 Data File : VU045316.D
 Acq On : 15 Oct 2021 11:35
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Oct 16 01:09:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U101321DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Oct 14 05:46:27 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 T	t-1,3-Dichloropropene	0.365	0.380	-4.1	99	0.00
51 T	cis-1,3-Dichloropropene	0.427	0.444	-4.0	99	0.00
52 RT	1,1,2-Trichloroethane	0.216	0.217	-0.5	98	0.00
53 t	1,3-Dichloropropane	0.402	0.401	0.2	100	0.00
54 t	2-Hexanone	0.158	0.167	-5.7	102	0.00
55 t	Dibromochloromethane	0.227	0.236	-4.0	99	0.00
56 T	1,2-Dibromoethane	0.203	0.202	0.5	96	0.00
57 S	4-Bromofluorobenzene	0.367	0.394	-7.4	102	0.00
58 RT	Tetrachloroethene	0.255	0.261	-2.4	103	0.00
59 Rt	Chlorobenzene	0.741	0.755	-1.9	99	0.00
60 T	1,1,1,2-Tetrachloroethane	0.246	0.258	-4.9	101	0.00
61 t	Pentachloroethane	0.179	0.190	-6.1	98	0.00
62 t	Hexachloroethane	0.188	0.204	-8.5	101	0.00
63 Rt	Ethyl Benzene	1.331	1.413	-6.2	100	0.00
64 RT	m/p-Xylenes	0.505	0.540	-6.9	100	0.00
65 RT	o-Xylene	0.493	0.516	-4.7	100	0.00
66 RT	Styrene	0.807	0.883	-9.4	101	0.00
67 t	Bromoform	0.119	0.123	-3.4	97	0.00
68 S	1,2-Dichlorobenzene-d4	0.344	0.325	5.5	89	0.00
69 T	Isopropylbenzene	1.296	1.375	-6.1	99	0.00
70 T	1,1,2,2-Tetrachloroethane	0.277	0.273	1.4	96	0.00
71 T	1,2,3-Trichloropropane	0.228	0.198	13.2	93	0.00
72 t	Bromobenzene	0.286	0.294	-2.8	99	0.00
73 t	n-propylbenzene	0.339	0.362	-6.8	99	0.00
74 t	2-Chlorotoluene	0.293	0.303	-3.4	98	0.00
75 t	1,3,5-Trimethylbenzene	1.081	1.169	-8.1	100	0.00
76 t	4-Chlorotoluene	0.299	0.318	-6.4	100	0.00
77 t	tert-Butylbenzene	1.049	1.094	-4.3	98	0.00
78 t	1,2,4-Trimethylbenzene	1.104	1.183	-7.2	98	0.00
79 t	sec-Butylbenzene	1.447	1.524	-5.3	99	0.00
80	Nitrobenzene	0.008	0.007	12.5	82	0.00
81 t	p-Isopropyltoluene	1.160	1.242	-7.1	98	0.00
82 t	1,3-Dichlorobenzene	0.572	0.580	-1.4	98	0.00
83 Rt	1,4-Dichlorobenzene	0.565	0.574	-1.6	98	0.00
84 t	n-Butylbenzene	1.106	1.211	-9.5	98	0.00
85 Rt	1,2-Dichlorobenzene	0.543	0.545	-0.4	97	0.00
86 t	1,2-Dibromo-3-Chloropropane	0.046	0.045	2.2	92	0.00
87 Rt	1,2,4-Trichlorobenzene	0.335	0.347	-3.6	95	0.00
88 t	Hexachlorobutadiene	0.157	0.164	-4.5	99	0.00
89 t	Naphthalene	0.658	0.639	2.9	86	0.00
90 t	1,2,3-Trichlorobenzene	0.311	0.317	-1.9	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0