

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101718\
 Data File : VU027671.D
 Acq On : 17 Oct 2018 12:11
 Operator : MD/SY
 Sample : J5463-09DL 20X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0JC0DL

Quant Time: Oct 19 06:58:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 19 06:41:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	118523	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	110905	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	54115	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	34428	3.66	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.20%
7) Chloroethane-d5	1.68	69	27275	3.85	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.00%
11) 1,1-Dichloroethene-d2	2.28	63	47559	2.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	56.80%#
20) 2-Butanone-d5	4.22	46	103095	38.49	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	76.98%
24) Chloroform-d	4.65	84	56135	3.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.40%
26) 1,2-Dichloroethane-d4	5.32	65	31142	3.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.60%
32) Benzene-d6	5.34	84	120386	3.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.60%
36) 1,2-Dichloropropane-d6	6.33	67	41461	3.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.60%
41) Toluene-d8	7.57	98	113181	3.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.40%
43) trans-1,3-Dichloropropene-	7.85	79	17289	3.72	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.40%
46) 2-Hexanone-d5	8.32	63	88845	38.99	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.98%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	30321	3.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	73.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	41184	4.06	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	81.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101718\
Data File : VU027671.D
Acq On : 17 Oct 2018 12:11
Operator : MD/SY
Sample : J5463-09DL 20X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0JC0DL

Quant Time: Oct 19 06:58:13 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 19 06:41:15 2018
Response via : Initial Calibration

