

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101718\
 Data File : VU027672.D
 Acq On : 17 Oct 2018 12:37
 Operator : MD/SY
 Sample : J5463-10DL 10X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0JC1DL

Quant Time: Oct 19 07:01:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 19 06:41:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	280319	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	264949	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	132240	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	96	0.00	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	0.00%#
7) Chloroethane-d5	1.69	69	126	0.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	0.20%#
11) 1,1-Dichloroethene-d2	2.29	63	926	0.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	0.40%#
20) 2-Butanone-d5	4.24	46	95	0.01	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	0.02%#
24) Chloroform-d	4.67	84	241	0.01	ug/L	0.01
Spiked Amount	5.000	Range	70 - 125	Recovery	=	0.20%#
26) 1,2-Dichloroethane-d4	5.31	65	475	0.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	0.40%#
32) Benzene-d6	5.34	84	770	0.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	0.20%#
36) 1,2-Dichloropropane-d6	6.34	67	200	0.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	0.20%#
41) Toluene-d8	7.57	98	1078	0.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	0.40%#
43) trans-1,3-Dichloropropene-	7.86	79	43	0.00	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	0.00%#
46) 2-Hexanone-d5	8.34	63	243	0.04	ug/L	0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	0.08%#
57) 1,1,2,2-Tetrachloroethane-	0.00	84	0d	0.00	ug/L	
Spiked Amount	5.000	Range	65 - 120	Recovery	=	0.00%#
64) 1,2-Dichlorobenzene-d4	11.86	152	1044	0.04	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	0.80%#

Target Compounds					Ovalue
22) cis-1,2-Dichloroethene	4.23	96	9808	0.439 ug/L #	98
34) Trichloroethene	6.19	95	4762	0.212 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101718\
Data File : VU027672.D
Acq On : 17 Oct 2018 12:37
Operator : MD/SY
Sample : J5463-10DL 10X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0JC1DL

Quant Time: Oct 19 07:01:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 19 06:41:15 2018
Response via : Initial Calibration

