

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101719\
 Data File : VU035214.D
 Acq On : 17 Oct 2019 16:47
 Operator : JC/SP
 Sample : K5318-01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CG0

Quant Time: Oct 18 04:30:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 18 03:45:16 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	396460	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	373796	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	157066	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	117035	4.84	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.80%
7) Chloroethane-d5	1.93	69	104227	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.40%
11) 1,1-Dichloroethene-d2	2.59	63	163449	3.66	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.20%
20) 2-Butanone-d5	4.72	46	296408	42.67	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.34%
24) Chloroform-d	5.11	84	206700	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.60%
26) 1,2-Dichloroethane-d4	5.75	65	119595	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.80%
32) Benzene-d6	5.77	84	429951	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.60%
36) 1,2-Dichloropropane-d6	6.74	67	135821	5.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.00%
41) Toluene-d8	7.93	98	410201	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.40%
43) trans-1,3-Dichloropropene-	8.22	79	62031	4.71	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.20%
46) 2-Hexanone-d5	8.69	63	259852	44.73	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.46%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	113131	4.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.60%
64) 1,2-Dichlorobenzene-d4	12.22	152	140174	4.90	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.00%

Target Compounds					Ovalue
13) Acetone	2.68	43	18822	3.895 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101719\
Data File : VU035214.D
Acq On : 17 Oct 2019 16:47
Operator : JC/SP
Sample : K5318-01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CG0

Quant Time: Oct 18 04:30:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 18 03:45:16 2019
Response via : Initial Calibration

