


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU101720\
 Data File : VU040767.D
 Acq On : 17 Oct 2020 15:30
 Operator : SY/MD
 Sample : BFB23
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB23

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Wed Oct 21 03:13:40 2020



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2874

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.2	48234	PASS
75	95	30	80	54.1	123061	PASS
95	95	100	100	100.0	227434	PASS
96	95	5	9	6.4	14482	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	75.4	171456	PASS
175	174	5	9	7.4	12689	PASS
176	174	95	101	95.0	162901	PASS
177	176	5	9	6.8	11125	PASS

Scan 2874 (10.581 min): VU040767.D
BFB23

m/z	Abundance
39.90	180.0
43.90	320.0

Instrument :
MSVOA_U
ClientSampleId :
BFB23

m/z	Abundance
36.00	2478.0
37.10	12859.0
38.10	11872.0
39.10	4553.0
39.90	224.0
43.10	107.0
44.00	1931.0
45.00	2243.0
46.00	345.0
47.00	3105.0
48.00	1822.0
49.10	11175.0
50.10	47952.0
51.10	15412.0
52.10	588.0
53.10	205.0
55.00	718.0
56.10	4164.0
57.10	7182.0
58.00	374.0
60.10	2577.0
61.10	12045.0
62.10	12152.0
63.10	10574.0
64.10	1601.0
65.00	1952.0
66.10	290.0
67.00	597.0
68.10	25392.0
69.10	24808.0
70.10	2214.0
72.10	1374.0
73.10	11175.0
74.10	42448.0
75.10	122528.0
76.10	10770.0
77.10	1403.0
78.00	861.0
79.00	10268.0
80.00	2855.0
81.00	10449.0
81.90	2422.0
83.00	143.0
86.10	392.0
87.00	9204.0
88.10	8774.0
91.10	1081.0
92.10	6909.0
93.10	10434.0
94.10	28680.0
95.10	224576.0
96.10	14012.0
97.00	400.0
103.90	1358.0
104.90	348.0
106.00	1250.0
106.90	292.0
109.90	145.0
111.00	236.0
111.90	134.0
113.00	259.0
115.00	187.0
116.00	992.0
117.00	1602.0
118.00	1109.0
119.00	1590.0
123.90	126.0
126.00	104.0
128.00	969.0
129.00	364.0
129.90	871.0
131.00	330.0
135.00	540.0
136.00	100.0
137.00	450.0
139.90	173.0
141.00	2794.0
142.00	436.0
143.10	3194.0
144.00	222.0
145.00	974.0
146.00	431.0
146.90	102.0
147.90	452.0
148.90	202.0
150.00	166.0
152.10	154.0
153.10	179.0
154.00	135.0
155.00	549.0
157.10	295.0
158.10	107.0

Instrument :
MSVOA_U
ClientSampleId :
BFB23

158.90	285.0
161.00	227.0
171.10	451.0
172.10	1874.0
174.10	164160.0
175.10	12048.0
176.10	155200.0
177.10	10649.0
178.10	258.0

Instrument :
MSVOA_U
ClientSampleId :
BFB23

m/z	Abundance
36.00	2448.0
37.10	13091.0
38.10	12388.0
39.10	4580.0
40.00	275.0
43.00	138.0
44.00	2040.0
45.00	2333.0
46.10	296.0
47.10	3246.0
48.00	1738.0
49.10	11544.0
50.10	49832.0
51.10	15988.0
52.00	545.0
53.00	165.0
55.00	820.0
56.10	4227.0
57.10	7443.0
58.00	383.0
60.10	2860.0
61.00	12777.0
62.10	12878.0
63.10	10598.0
64.10	1433.0
65.00	1728.0
66.10	299.0
67.00	603.0
68.10	26496.0
69.10	25736.0
70.10	2116.0
72.10	1358.0
73.10	11842.0
74.10	44432.0
75.10	127136.0
76.10	11124.0
77.10	1621.0
78.00	821.0
79.00	10344.0
80.00	2957.0
81.00	10767.0
81.90	2424.0
83.00	199.0
86.20	406.0
87.10	9640.0
88.10	9238.0
91.00	1045.0
92.10	6963.0
93.10	11065.0
94.10	30792.0
95.10	235200.0
96.10	14986.0
97.00	495.0
102.90	119.0
104.00	1303.0
104.90	340.0
106.00	1292.0
106.90	316.0
110.00	185.0
110.90	300.0
111.90	188.0
113.00	257.0
115.00	252.0
116.00	1015.0
117.00	1624.0
118.00	1187.0
119.00	1581.0
123.90	130.0
126.00	121.0
128.00	958.0
129.00	411.0
129.90	865.0
131.10	304.0
135.00	572.0
135.90	100.0
137.00	525.0
139.90	150.0
141.00	3004.0
142.00	433.0
143.10	3246.0
144.10	229.0
145.00	860.0
146.00	421.0
146.90	113.0
147.90	531.0
148.90	283.0
150.10	229.0
152.00	152.0
153.10	168.0
154.00	115.0
155.00	546.0
157.10	349.0

Instrument :
MSVOA_U
ClientSampleId :
BFB23

158.10	121.0
158.90	281.0
161.00	262.0
171.10	362.0
172.10	1675.0
174.10	177536.0
175.10	13143.0
176.10	168128.0
177.10	11523.0
178.10	282.0

Instrument :
MSVOA_U
ClientSampleId :
BFB23

m/z	Abundance
36.00	2314.0
37.00	12287.0
38.00	11600.0
39.10	4376.0
40.00	323.0
43.00	137.0
44.00	1935.0
45.00	2172.0
46.20	237.0
47.10	2941.0
48.00	1506.0
49.10	11018.0
50.10	46920.0
51.10	14704.0
52.00	521.0
53.00	135.0
55.10	779.0
56.10	3742.0
57.10	6868.0
58.00	361.0
60.00	2662.0
61.00	12018.0
62.10	12179.0
63.10	9680.0
64.10	1245.0
65.00	1389.0
66.10	223.0
67.00	579.0
68.10	24728.0
69.10	23960.0
70.10	1819.0
72.10	1301.0
73.10	11359.0
74.10	42128.0
75.10	119520.0
76.10	10401.0
77.10	1697.0
78.00	716.0
79.00	9528.0
80.00	2668.0
81.00	9939.0
82.00	2166.0
83.00	240.0
86.20	404.0
87.00	9374.0
88.10	8859.0
91.00	980.0
92.10	6502.0
93.10	10366.0
94.10	29176.0
95.10	222528.0
96.10	14448.0
97.00	513.0
102.90	127.0
104.00	1210.0
104.90	323.0
106.00	1226.0
107.00	307.0
109.90	177.0
110.90	349.0
111.90	226.0
113.00	237.0
115.00	302.0
116.00	891.0
117.00	1617.0
117.90	1096.0
119.00	1492.0
123.90	126.0
126.00	118.0
128.00	896.0
128.90	415.0
129.90	814.0
131.10	276.0
135.00	526.0
136.90	538.0
140.10	143.0
141.00	2847.0
142.00	372.0
143.00	3049.0
144.10	210.0
145.00	694.0
146.00	321.0
146.90	128.0
147.90	524.0
148.90	269.0
150.10	266.0
151.90	116.0
153.00	160.0
155.00	477.0
157.10	393.0
159.00	234.0
161.00	313.0

Instrument :
MSVOA_U
ClientSampleId :
BFB23

170.90	257.0
172.00	1470.0
174.10	172672.0
175.10	12878.0
176.10	165376.0
177.10	11203.0
178.10	295.0

Instrument :
MSVOA_U
ClientSampleId :
BFB23