

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101818\
 Data File : VU027695.D
 Acq On : 18 Oct 2018 13:39
 Operator : MD/SY
 Sample : J5463-19 200X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0JD4

Quant Time: Oct 19 08:50:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 19 08:11:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	75819	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	69001	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	33466	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	23264	3.87	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.40%
7) Chloroethane-d5	1.68	69	18930	4.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.60%
11) 1,1-Dichloroethene-d2	2.28	63	32070	3.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.00%
20) 2-Butanone-d5	4.20	46	79799	46.58	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.16%
24) Chloroform-d	4.65	84	39136	4.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.40%
26) 1,2-Dichloroethane-d4	5.31	65	23960	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.00%
32) Benzene-d6	5.34	84	88420	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.00%
36) 1,2-Dichloropropane-d6	6.33	67	30929	4.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.40%
41) Toluene-d8	7.57	98	79051	4.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.00%
43) trans-1,3-Dichloropropene-	7.85	79	11139	3.85	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.00%
46) 2-Hexanone-d5	8.31	63	67948	47.93	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.86%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	22358	4.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.40%
64) 1,2-Dichlorobenzene-d4	11.86	152	28834	4.60	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.00%

Target Compounds					Ovalue
16) Methylene chloride	2.70	84	6276	1.121 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101818\
Data File : VU027695.D
Acq On : 18 Oct 2018 13:39
Operator : MD/SY
Sample : J5463-19 200X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0JD4

Quant Time: Oct 19 08:50:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 19 08:11:17 2018
Response via : Initial Calibration

