

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101819\
 Data File : VU035294.D
 Acq On : 19 Oct 2019 02:58
 Operator : JC/SP
 Sample : K5419-20
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Oct 19 06:54:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 19 06:13:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	310201	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	296376	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	118466	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	78661	4.16	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.20%
7) Chloroethane-d5	1.93	69	75224	4.35	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.00%
11) 1,1-Dichloroethene-d2	2.59	63	107928	3.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.80%
20) 2-Butanone-d5	4.71	46	269512	49.58	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.16%
24) Chloroform-d	5.11	84	161293	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.40%
26) 1,2-Dichloroethane-d4	5.75	65	92881	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.00%
32) Benzene-d6	5.77	84	306607	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.40%
36) 1,2-Dichloropropane-d6	6.74	67	103755	4.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.40%
41) Toluene-d8	7.93	98	270636	4.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.40%
43) trans-1,3-Dichloropropene-	8.22	79	41826	4.01	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.20%
46) 2-Hexanone-d5	8.68	63	212176	46.07	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.14%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	92540	4.82	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.40%
64) 1,2-Dichlorobenzene-d4	12.22	152	99221	4.60	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.00%


Target Compounds					Ovalue
16) Methylene chloride	3.09	84	3431	0.143 ug/L	90




(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101819\
Data File : VU035294.D
Acq On : 19 Oct 2019 02:58
Operator : JC/SP
Sample : K5419-20
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Oct 19 06:54:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 19 06:13:43 2019
Response via : Initial Calibration





#16

Methvlene chloride

Concen: 0.143 ug/L

RT: 3.09 min Scan# 543

Delta R.T. 0.00 min

Lab File: VU035294.D

Acq: 19 Oct 2019 02:58

Instrument :

MSVOA_U

ClientSampleId :

VHBLK01

Tot Ion: 84 Resp: 3431

Ion Ratio Lower Upper

84 100

86 62.3 45.4 84.2

49 109.3 86.8 161.2

