

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU101919\
 Data File : VU035307.D
 Acq On : 19 Oct 2019 10:21
 Operator : JC/SP
 Sample : VU1019WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK49

Quant Time: Oct 21 06:00:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM101919WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Oct 21 05:38:49 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	650079	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	622905	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	207391	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	267469	47.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.66%
7) Chloroethane-d5	1.93	69	230007	49.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.02%
11) 1,1-Dichloroethene-d2	2.60	63	345324	37.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.84%
21) 2-Butanone-d5	4.70	46	384946	103.85	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.85%
24) Chloroform-d	5.11	84	436323	48.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.56%
26) 1,2-Dichloroethane-d4	5.75	65	295392	49.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.72%
32) Benzene-d6	5.77	84	912909	50.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.64%
36) 1,2-Dichloropropane-d6	6.74	67	295904	51.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.10%
41) Toluene-d8	7.93	98	847219	49.93	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.86%
43) trans-1,3-Dichloropropene-	8.22	79	144241	48.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.02%
47) 2-Hexanone-d5	8.68	63	256551	100.59	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.59%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	404204	47.37	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.74%
64) 1,2-Dichlorobenzene-d4	12.22	152	235760	49.03	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.06%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU101919\
Data File : VU035307.D
Acq On : 19 Oct 2019 10:21
Operator : JC/SP
Sample : VU1019WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK49

Quant Time: Oct 21 06:00:09 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM101919WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Oct 21 05:38:49 2019
Response via : Initial Calibration

