

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U101920\
 Data File : VU040794.D
 Acq On : 19 Oct 2020 11:04
 Operator : SY/MD
 Sample : L4354-22
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETZ42

Quant Time: Oct 20 01:51:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 20 01:28:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	116867	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	121001	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	48760	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	34676	3.58	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.60%
7) Chloroethane-d5	1.92	69	31938	4.28	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.60%
11) 1,1-Dichloroethene-d2	2.58	63	53293	2.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	56.40%#
20) 2-Butanone-d5	4.66	46	183388	59.18	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.36%
24) Chloroform-d	5.08	84	81997	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
26) 1,2-Dichloroethane-d4	5.72	65	51096	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.80%
32) Benzene-d6	5.74	84	144254	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.40%
36) 1,2-Dichloropropane-d6	6.70	67	51015	4.71	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.20%
41) Toluene-d8	7.91	98	109395	3.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.80%
43) trans-1,3-Dichloropropene-	8.19	79	16866	3.67	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.40%
46) 2-Hexanone-d5	8.64	63	110629	46.77	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.54%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	48350	4.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	46562	5.15	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.00%

Target Compounds

					Ovalue
3) Chloromethane	1.53	50	853	0.086 ug/L	95
5) Vinyl chloride	1.61	62	160426	15.384 ug/L	96
12) 1,1-Dichloroethene	2.59	96	2150	0.258 ug/L #	1
18) trans-1,2-Dichloroethene	3.37	96	5848	0.670 ug/L	87
22) cis-1,2-Dichloroethene	4.69	96	994244	105.624 ug/L	98
34) Trichloroethene	6.56	95	12029	1.194 ug/L	94
47) Tetrachloroethene	8.56	164	6613	0.920 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101920\
Data File : VU040794.D
Acq On : 19 Oct 2020 11:04
Operator : SY/MD
Sample : L4354-22
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETZ42

Quant Time: Oct 20 01:51:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 20 01:28:45 2020
Response via : Initial Calibration

