

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U101920\
 Data File : VU040824.D
 Acq On : 19 Oct 2020 23:28
 Operator : SY/MD
 Sample : L4420-09
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AJ9

Quant Time: Oct 20 06:00:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 20 05:12:18 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	111739	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	116639	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	47864	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	26791	2.89	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	57.80%
7) Chloroethane-d5	1.92	69	29224	4.09	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.80%
11) 1,1-Dichloroethene-d2	2.58	63	45801	2.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	50.60%#
20) 2-Butanone-d5	4.66	46	168387	56.83	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.66%
24) Chloroform-d	5.08	84	79107	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.40%
26) 1,2-Dichloroethane-d4	5.72	65	50208	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.40%
32) Benzene-d6	5.74	84	130790	4.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.20%
36) 1,2-Dichloropropane-d6	6.70	67	49622	4.75	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.00%
41) Toluene-d8	7.91	98	97139	3.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.00%#
43) trans-1,3-Dichloropropene-	8.19	79	18227	4.12	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.40%
46) 2-Hexanone-d5	8.64	63	110904	48.64	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.28%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	47693	5.09	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	45742	5.15	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.00%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.59	96	1817	0.228 ug/L	# 1
22) cis-1,2-Dichloroethene	4.69	96	2927	0.325 ug/L	95
25) Chloroform	5.10	83	3644	0.207 ug/L	82
31) Carbon tetrachloride	5.54	117	3391	0.253 ug/L	98
34) Trichloroethene	6.55	95	11743	1.209 ug/L	94
47) Tetrachloroethene	8.56	164	6930	1.000 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101920\
Data File : VU040824.D
Acq On : 19 Oct 2020 23:28
Operator : SY/MD
Sample : L4420-09
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AJ9

Quant Time: Oct 20 06:00:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 20 05:12:18 2020
Response via : Initial Calibration

