

Method Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\  
 Method File : SFAMULM101921WMA.M  
 Title : VOC Analysis  
 Last Update : Wed Oct 20 02:57:45 2021  
 Response Via : Initial Calibration

## Calibration Files

5 =VU045325.D 10 =VU045326.D 50 =VU045327.D 100 =VU045328.D 200 =VU045329.D

| Compound |                     | 5              | 10    | 50    | 100   | 200   | Avg   | %RSD  |
|----------|---------------------|----------------|-------|-------|-------|-------|-------|-------|
| -----    |                     |                |       |       |       |       |       |       |
| 1) I     | 1,4-Difluorobenzene | -----ISTD----- |       |       |       |       |       |       |
| 2) T     | Dichlorodifluoro... | 0.396          | 0.389 | 0.346 | 0.345 | 0.346 | 0.365 | 7.06  |
| 3) T     | Chloromethane       | 0.473          | 0.431 | 0.397 | 0.385 | 0.375 | 0.412 | 9.69  |
| 4) S     | Vinyl Chloride-d3   | 0.225          | 0.223 | 0.215 | 0.214 | 0.212 | 0.218 | 2.80  |
| 5) T     | Vinyl chloride      | 0.509          | 0.478 | 0.435 | 0.429 | 0.430 | 0.456 | 7.84  |
| 6) T     | Bromomethane        | 0.260          | 0.246 | 0.228 | 0.238 | 0.235 | 0.241 | 5.02  |
| 7) S     | Chloroethane-d5     | 0.222          | 0.205 | 0.206 | 0.206 | 0.202 | 0.208 | 3.69  |
| 8) T     | Chloroethane        | 0.293          | 0.316 | 0.274 | 0.276 | 0.274 | 0.287 | 6.37  |
| 9) T     | Trichlorofluorom... | 0.632          | 0.575 | 0.520 | 0.524 | 0.525 | 0.555 | 8.76  |
| 10) T    | 1,1,2-Trichloro-... | 0.377          | 0.374 | 0.335 | 0.340 | 0.338 | 0.353 | 5.87  |
| 11) S    | 1,1-Dichloroethe... | 0.453          | 0.437 | 0.418 | 0.427 | 0.433 | 0.434 | 2.96  |
| 12) T    | 1,1-Dichloroethene  | 0.350          | 0.332 | 0.311 | 0.313 | 0.319 | 0.325 | 5.02  |
| 13) T    | Acetone             | 0.475          | 0.423 | 0.362 | 0.307 | 0.326 | 0.379 | 18.45 |
| 14) T    | Carbon disulfide    | 0.970          | 0.909 | 0.855 | 0.866 | 0.869 | 0.894 | 5.29  |
| 15) T    | Methyl Acetate      | 0.515          | 0.489 | 0.447 | 0.449 | 0.447 | 0.469 | 6.64  |
| 16) T    | Methylene chloride  | 0.488          | 0.444 | 0.390 | 0.385 | 0.387 | 0.419 | 10.96 |
| 17) T    | trans-1,2-Dichlo... | 0.396          | 0.349 | 0.329 | 0.334 | 0.336 | 0.349 | 7.81  |
| 18) T    | Methyl tert-buty... | 1.255          | 1.212 | 1.189 | 1.204 | 1.223 | 1.216 | 2.05  |
| 19) T    | 1,1-Dichloroethane  | 0.787          | 0.739 | 0.684 | 0.684 | 0.675 | 0.714 | 6.74  |
| 20) T    | cis-1,2-Dichloro... | 0.423          | 0.403 | 0.386 | 0.394 | 0.392 | 0.400 | 3.69  |
| 21) S    | 2-Butanone-d5       | 0.315          | 0.311 | 0.329 | 0.334 | 0.336 | 0.325 | 3.45  |
| 22) T    | 2-Butanone          | 0.474          | 0.430 | 0.412 | 0.382 | 0.390 | 0.417 | 8.83  |
| 23) T    | Bromochloromethane  | 0.214          | 0.206 | 0.189 | 0.187 | 0.189 | 0.197 | 6.17  |
| 24) S    | Chloroform-d        | 0.580          | 0.571 | 0.577 | 0.586 | 0.583 | 0.579 | 0.95  |
| 25) T    | Chloroform          | 0.810          | 0.753 | 0.702 | 0.687 | 0.682 | 0.727 | 7.50  |
| 26) S    | 1,2-Dichloroetha... | 0.424          | 0.397 | 0.390 | 0.386 | 0.382 | 0.396 | 4.23  |
| 27) T    | 1,2-Dichloroethane  | 0.646          | 0.591 | 0.557 | 0.558 | 0.550 | 0.580 | 6.86  |
|          |                     |                |       |       |       |       |       |       |
| 28) I    | Chlorobenzene-d5    | -----ISTD----- |       |       |       |       |       |       |
| 29) T    | Cyclohexane         | 0.591          | 0.587 | 0.583 | 0.601 | 0.607 | 0.594 | 1.72  |
| 30) T    | 1,1,1-Trichloroe... | 0.649          | 0.591 | 0.567 | 0.570 | 0.573 | 0.590 | 5.82  |
| 31) T    | Carbon tetrachlo... | 0.485          | 0.462 | 0.440 | 0.444 | 0.447 | 0.456 | 4.04  |
| 32) S    | Benzene-d6          | 0.947          | 0.921 | 0.951 | 0.949 | 0.933 | 0.940 | 1.37  |
| 33) T    | Benzene             | 1.670          | 1.586 | 1.517 | 1.507 | 1.502 | 1.557 | 4.61  |
| 34) T    | Trichloroethene     | 0.414          | 0.388 | 0.373 | 0.372 | 0.373 | 0.384 | 4.75  |
| 35) T    | Methylcyclohexane   | 0.607          | 0.592 | 0.585 | 0.610 | 0.621 | 0.603 | 2.41  |
| 36) S    | 1,2-Dichloroprop... | 0.365          | 0.361 | 0.366 | 0.368 | 0.367 | 0.366 | 0.68  |
| 37) T    | 1,2-Dichloropropane | 0.455          | 0.441 | 0.410 | 0.407 | 0.404 | 0.423 | 5.45  |
| 38) T    | Bromodichloromet... | 0.587          | 0.561 | 0.520 | 0.531 | 0.538 | 0.548 | 4.91  |
| 39) T    | cis-1,3-Dichloro... | 0.629          | 0.610 | 0.626 | 0.643 | 0.662 | 0.634 | 3.09  |
| 40) T    | 4-Methyl-2-penta... | 0.636          | 0.644 | 0.636 | 0.646 | 0.668 | 0.646 | 2.04  |
| 41) S    | Toluene-d8          | 0.791          | 0.787 | 0.801 | 0.819 | 0.813 | 0.802 | 1.72  |
| 42) T    | Toluene             | 1.685          | 1.642 | 1.582 | 1.602 | 1.609 | 1.624 | 2.49  |
| 43) S    | trans-1,3-Dichlo... | 0.171          | 0.174 | 0.186 | 0.194 | 0.203 | 0.186 | 7.15  |
| 44) T    | trans-1,3-Dichlo... | 0.598          | 0.579 | 0.601 | 0.615 | 0.637 | 0.606 | 3.59  |
| 45) T    | 1,1,2-Trichloroe... | 0.431          | 0.428 | 0.393 | 0.387 | 0.390 | 0.406 | 5.35  |
| 46) T    | Tetrachloroethene   | 0.283          | 0.269 | 0.254 | 0.261 | 0.263 | 0.266 | 4.12  |
| 47) S    | 2-Hexanone-d5       | 0.183          | 0.212 | 0.238 | 0.249 | 0.266 | 0.229 | 14.22 |
| 48) T    | 2-Hexanone          | 0.556          | 0.537 | 0.551 | 0.528 | 0.553 | 0.545 | 2.22  |
| 49) T    | Dibromochloromet... | 0.412          | 0.389 | 0.380 | 0.391 | 0.404 | 0.395 | 3.19  |
| 50) T    | 1,2-Dibromoethane   | 0.447          | 0.423 | 0.402 | 0.404 | 0.412 | 0.418 | 4.40  |
| 51) T    | Chlorobenzene       | 1.102          | 1.046 | 0.980 | 0.995 | 1.019 | 1.028 | 4.68  |
| 52) T    | Ethylbenzene        | 1.764          | 1.733 | 1.720 | 1.777 | 1.816 | 1.762 | 2.14  |
| 53) T    | m,p-Xylene          | 0.657          | 0.651 | 0.660 | 0.683 | 0.697 | 0.669 | 2.94  |
| 54) T    | o-xylene            | 0.674          | 0.656 | 0.662 | 0.672 | 0.694 | 0.672 | 2.14  |
| 55) T    | Styrene             | 1.055          | 1.084 | 1.142 | 1.177 | 1.226 | 1.137 | 6.06  |
| 56) S    | 1,1,2,2-Tetrachl... | 0.696          | 0.681 | 0.671 | 0.692 | 0.715 | 0.691 | 2.42  |

Method Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\  
 Method File : SFAMULM101921WMA.M

|     |   |                       |                |       |       |       |       |       |       |
|-----|---|-----------------------|----------------|-------|-------|-------|-------|-------|-------|
| 57) | T | 1,1,2,2-Tetrachl...   | 0.759          | 0.732 | 0.684 | 0.694 | 0.715 | 0.717 | 4.19  |
| 58) | I | 1,4-Dichlorobenzen... | -----ISTD----- |       |       |       |       |       |       |
| 59) | T | Bromoform             | 0.556          | 0.550 | 0.519 | 0.543 | 0.572 | 0.548 | 3.57  |
| 60) | T | 1,2,3-Trichlorop...   | 1.288          | 1.218 | 1.083 | 1.068 | 1.083 | 1.148 | 8.64  |
| 61) | T | Isopropylbenzene      | 3.438          | 3.474 | 3.302 | 3.393 | 3.464 | 3.414 | 2.05  |
| 62) | T | 1,3,5-Trimethylb...   | 2.699          | 2.670 | 2.821 | 2.926 | 3.008 | 2.825 | 5.11  |
| 63) | T | 1,2,4-Trimethylb...   | 2.682          | 2.710 | 2.863 | 2.973 | 3.073 | 2.860 | 5.86  |
| 64) | T | 1,3-Dichlorobenzene   | 1.669          | 1.570 | 1.471 | 1.487 | 1.534 | 1.546 | 5.11  |
| 65) | T | 1,4-Dichlorobenzene   | 1.711          | 1.616 | 1.476 | 1.497 | 1.541 | 1.568 | 6.14  |
| 66) | S | 1,2-Dichlorobenz...   | 0.882          | 0.865 | 0.842 | 0.862 | 0.887 | 0.868 | 2.06  |
| 67) | T | 1,2-Dichlorobenzene   | 1.681          | 1.640 | 1.472 | 1.495 | 1.526 | 1.563 | 5.91  |
| 68) | T | 1,2-Dibromo-3-ch...   | 0.326          | 0.316 | 0.304 | 0.319 | 0.333 | 0.320 | 3.47  |
| 69) |   | 1,3,5-Trichlorob...   | 1.062          | 1.038 | 1.011 | 1.078 | 1.131 | 1.064 | 4.23  |
| 70) | T | 1,2,4-trichlorob...   | 0.777          | 0.830 | 0.867 | 0.958 | 1.027 | 0.892 | 11.27 |
| 71) |   | Naphthalene           | 2.117          | 2.503 | 3.093 | 3.510 | 3.682 | 2.981 | 22.24 |
| 72) | T | 1,2,3-Trichlorob...   | 0.762          | 0.837 | 0.892 | 0.963 | 1.000 | 0.891 | 10.76 |

(#) = Out of Range