

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101921\
 Data File : VU045331.D
 Acq On : 19 Oct 2021 15:48
 Operator : SY/MD
 Sample : VU1019MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK025

Quant Time: Oct 20 03:08:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM101921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 20 02:57:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	294503	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	300148	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	133251	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	58024	45.228	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.460%
7) Chloroethane-d5	1.874	69	22048	17.984	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	35.960%#
11) 1,1-Dichloroethene-d2	2.552	63	71774	28.097	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	56.200%#
21) 2-Butanone-d5	4.629	46	203137	106.078	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.080%
24) Chloroform-d	5.063	84	160776	47.121	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.240%
26) 1,2-Dichloroethane-d4	5.703	65	115584	49.559	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.120%
32) Benzene-d6	5.726	84	275689	48.834	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.660%
36) 1,2-Dichloropropane-d6	6.694	67	109767	50.019	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.040%
41) Toluene-d8	7.899	98	225771	46.887	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.780%
43) trans-1,3-Dichloroprop...	8.182	79	55204	49.559	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.120%
47) 2-Hexanone-d5	8.639	63	168402	122.303	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	122.300%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	241182	58.135	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	116.260%
66) 1,2-Dichlorobenzene-d4	12.198	152	128563	55.608	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101921\
Data File : VU045331.D
Acq On : 19 Oct 2021 15:48
Operator : SY/MD
Sample : VU1019MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK025

Quant Time: Oct 20 03:08:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM101921WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 20 02:57:45 2021
Response via : Initial Calibration

