

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101923\
 Data File : VU055809.D
 Acq On : 19 Oct 2023 10:51
 Operator : MD/SY
 Sample : VU1019WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK13

Quant Time: Oct 20 01:37:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U101823DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Oct 19 03:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.113	96	73373	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.634	95	24401	0.843	ug/l	0.00
Spiked Amount	1.000		Recovery	=	84.000%	
68) 1,2-Dichlorobenzene-d4	12.193	152	21184	0.790	ug/l	0.00
Spiked Amount	1.000		Recovery	=	79.000%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101923\
Data File : VU055809.D
Acq On : 19 Oct 2023 10:51
Operator : MD/SY
Sample : VU1019WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK13

Quant Time: Oct 20 01:37:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U101823DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Oct 19 03:10:04 2023
Response via : Initial Calibration

