

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U102020\
 Data File : VU040843.D
 Acq On : 20 Oct 2020 13:15
 Operator : SY/MD
 Sample : L4354-13
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETZ24

Quant Time: Oct 21 00:15:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 20 23:49:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	124452	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	133507	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	55366	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	42748	4.14	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.80%
7) Chloroethane-d5	1.92	69	37547	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.40%
11) 1,1-Dichloroethene-d2	2.58	63	67316	3.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.80%
20) 2-Butanone-d5	4.65	46	163199	49.46	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.92%
24) Chloroform-d	5.08	84	86649	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
26) 1,2-Dichloroethane-d4	5.72	65	53428	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.00%
32) Benzene-d6	5.74	84	158384	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.00%
36) 1,2-Dichloropropane-d6	6.70	67	53796	4.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.00%
41) Toluene-d8	7.91	98	125555	3.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.80%
43) trans-1,3-Dichloropropene-	8.19	79	20278	4.00	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.00%
46) 2-Hexanone-d5	8.64	63	118200	45.29	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.58%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	49184	4.59	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	52668	5.13	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.60%

Target Compounds					Ovalue
13) Acetone	2.66	43	2623	1.025 ug/L	99
52) Ethylbenzene	9.57	91	2066	0.044 ug/L	95
53) m,p-Xylene	9.69	106	2764	0.160 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102020\
Data File : VU040843.D
Acq On : 20 Oct 2020 13:15
Operator : SY/MD
Sample : L4354-13
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETZ24

Quant Time: Oct 21 00:15:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 20 23:49:45 2020
Response via : Initial Calibration

