


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU102020\
 Data File : VU040839.D
 Acq On : 20 Oct 2020 10:26
 Operator : SY/MD
 Sample : BFB25
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB25

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Wed Oct 21 08:06:50 2020



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2875

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.6	32888	PASS
75	95	30	80	55.2	83904	PASS
95	95	100	100	100.0	152021	PASS
96	95	5	9	6.3	9566	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	74.9	113834	PASS
175	174	5	9	7.6	8639	PASS
176	174	95	101	95.0	108170	PASS
177	176	5	9	6.5	7062	PASS

m/z	Abundance
36.00	1635.0
37.10	8241.0
38.00	8374.0
39.10	3623.0
40.00	285.0
42.90	103.0
44.00	1506.0
45.00	1544.0
46.00	229.0
47.10	2012.0
48.00	941.0
49.00	6955.0
50.10	32168.0
51.10	10427.0
52.10	501.0
53.00	282.0
55.00	500.0
56.10	2980.0
57.00	4306.0
58.10	167.0
60.00	1655.0
61.10	8420.0
62.10	8232.0
63.10	6795.0
64.10	1191.0
65.10	1569.0
66.00	194.0
67.10	408.0
68.10	16952.0
69.10	16536.0
70.00	1277.0
72.10	1123.0
73.00	7158.0
74.10	27432.0
75.10	83024.0
76.10	6929.0
77.00	939.0
78.20	715.0
79.00	6935.0
80.00	2055.0
81.00	7309.0
82.00	1841.0
83.00	137.0
85.90	267.0
87.00	6030.0
88.00	5599.0
91.00	794.0
92.10	5032.0
93.10	6793.0
94.10	19248.0
95.10	148928.0
96.10	9572.0
97.10	320.0
104.00	988.0
104.90	334.0
106.00	901.0
107.00	165.0
108.00	106.0
110.90	283.0
112.00	105.0
115.10	168.0
115.90	607.0
117.00	1444.0
117.90	734.0
119.00	1166.0
128.00	500.0
129.00	267.0
130.00	545.0
130.90	344.0
134.90	276.0
136.90	293.0
140.00	189.0
141.00	1763.0
142.00	276.0
143.00	2716.0
144.00	153.0
145.00	735.0
145.90	183.0
148.00	431.0
155.00	338.0
157.10	210.0
158.90	133.0
161.00	247.0
171.10	380.0
172.10	1690.0
174.10	108112.0
175.10	8460.0
176.10	102816.0
177.10	6915.0
178.00	184.0

Instrument :
MSVOA_U
ClientSampleId :
BFB25

m/z	Abundance
36.00	1694.0
37.10	8819.0
38.00	8654.0
39.00	3648.0
40.00	314.0
44.00	1470.0
45.00	1606.0
46.00	234.0
47.10	2217.0
47.90	1033.0
49.00	7190.0
50.10	34144.0
51.10	10285.0
52.10	488.0
53.00	260.0
55.00	512.0
56.10	2928.0
57.10	4611.0
58.10	186.0
60.00	1840.0
61.10	8879.0
62.10	8480.0
63.10	7273.0
64.10	1170.0
65.10	1639.0
66.10	172.0
67.10	451.0
68.10	17184.0
69.10	17440.0
70.00	1421.0
72.10	1240.0
73.00	7518.0
74.10	28744.0
75.10	86912.0
76.10	7158.0
77.00	1108.0
77.70	559.0
79.00	6994.0
80.00	2057.0
81.00	7636.0
82.00	1934.0
83.00	164.0
86.00	294.0
87.10	6244.0
88.00	5863.0
91.00	880.0
92.10	4927.0
93.10	7180.0
94.10	20464.0
95.10	157248.0
96.10	9850.0
97.00	312.0
102.70	102.0
104.00	953.0
104.90	270.0
106.00	859.0
107.00	163.0
110.90	297.0
113.00	111.0
115.00	215.0
116.00	627.0
117.00	1422.0
117.90	730.0
119.00	1179.0
123.80	110.0
128.00	514.0
128.90	232.0
130.00	526.0
130.90	361.0
135.00	269.0
136.90	322.0
140.00	185.0
141.00	1912.0
142.00	231.0
143.00	2765.0
144.00	144.0
145.00	642.0
146.00	198.0
148.00	540.0
149.90	104.0
155.00	349.0
156.00	103.0
157.10	263.0
158.90	152.0
161.00	233.0
171.00	393.0
172.10	1598.0
174.10	117880.0
175.10	8910.0
176.10	111656.0
177.10	7345.0
178.00	199.0

Instrument :
MSVOA_U
ClientSampleId :
BFB25

m/z	Abundance
36.00	1519.0
37.00	8593.0
38.00	7665.0
39.00	3231.0
40.00	311.0
44.00	1295.0
45.00	1395.0
46.00	214.0
47.00	2190.0
47.90	1049.0
49.10	6613.0
50.10	32352.0
51.10	9437.0
52.10	436.0
53.10	230.0
55.00	499.0
56.10	2560.0
57.10	4562.0
58.10	173.0
60.00	1776.0
61.10	8354.0
62.10	7888.0
63.10	6984.0
64.10	1026.0
65.10	1511.0
66.10	154.0
67.10	388.0
68.10	15707.0
69.10	16285.0
70.00	1392.0
72.10	1134.0
73.10	7589.0
74.10	27296.0
75.10	81776.0
76.10	6670.0
77.00	1005.0
77.80	502.0
79.00	6627.0
80.00	1950.0
81.00	7259.0
82.00	1726.0
83.00	191.0
86.00	335.0
87.10	5994.0
88.00	5869.0
91.00	783.0
92.10	4529.0
93.10	6982.0
94.10	19408.0
95.10	149888.0
96.10	9277.0
97.00	248.0
102.80	123.0
104.00	805.0
104.90	219.0
106.00	738.0
106.90	174.0
110.90	227.0
112.00	121.0
112.90	120.0
115.00	238.0
116.00	620.0
117.00	1287.0
118.00	709.0
119.00	1078.0
128.00	525.0
128.90	213.0
130.00	492.0
131.00	312.0
135.00	267.0
137.00	325.0
140.10	176.0
141.00	1908.0
142.00	171.0
143.10	2560.0
144.10	104.0
145.00	550.0
146.00	263.0
148.00	494.0
155.00	333.0
156.00	126.0
157.00	251.0
158.90	148.0
161.00	171.0
171.00	344.0
172.10	1447.0
174.10	115512.0
175.10	8548.0
176.10	110040.0
177.10	6928.0
178.00	198.0

Instrument :
MSVOA_U
ClientSampleId :
BFB25

Scan 2875 (10.584 min): VU040839.D
BFB25

m/z	Abundance
39.90	189.0
43.90	295.0

Instrument :
MSVOA_U
ClientSampleId :
BFB25