

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102120\
 Data File : VU040876.D
 Acq On : 21 Oct 2020 16:21
 Operator : SY/MD
 Sample : VIBLK78
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK78

Quant Time: Oct 22 01:29:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 22 00:42:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	229190	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	241639	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	105598	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	74326	4.73	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.60%
7) Chloroethane-d5	1.92	69	67759	4.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.00%
11) 1,1-Dichloroethene-d2	2.58	63	113365	3.39	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.80%
20) 2-Butanone-d5	4.65	46	345775	61.98	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.96%
24) Chloroform-d	5.08	84	162796	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
26) 1,2-Dichloroethane-d4	5.72	65	101059	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.60%
32) Benzene-d6	5.74	84	290267	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.80%
36) 1,2-Dichloropropane-d6	6.70	67	102836	4.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.60%
41) Toluene-d8	7.91	98	228258	4.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.20%
43) trans-1,3-Dichloropropene-	8.19	79	40294	4.68	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.60%
46) 2-Hexanone-d5	8.64	63	253109	58.49	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	116.98%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	95149	5.20	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	100102	5.39	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.80%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	7268	0.352 ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102120\
Data File : VU040876.D
Acq On : 21 Oct 2020 16:21
Operator : SY/MD
Sample : VIBLK78
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK78

Quant Time: Oct 22 01:29:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 22 00:42:50 2020
Response via : Initial Calibration

