

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102120\
 Data File : VU040900.D
 Acq On : 22 Oct 2020 11:50
 Operator : SY/MD
 Sample : VU1021WBL02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK07

Quant Time: Oct 23 00:07:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 22 23:17:57 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	208654	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	227016	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	92923	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	67474	4.72	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.40%
7) Chloroethane-d5	1.92	69	63986	5.13	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.60%
11) 1,1-Dichloroethene-d2	2.58	63	103685	3.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.20%
20) 2-Butanone-d5	4.65	46	326157	64.22	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	128.44%
24) Chloroform-d	5.08	84	154873	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.60%
26) 1,2-Dichloroethane-d4	5.72	65	99911	5.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.60%
32) Benzene-d6	5.74	84	271354	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
36) 1,2-Dichloropropane-d6	6.70	67	94935	4.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.00%
41) Toluene-d8	7.91	98	201196	3.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.20%
43) trans-1,3-Dichloropropene-	8.18	79	35774	4.42	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.40%
46) 2-Hexanone-d5	8.64	63	227479	55.96	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.92%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	90453	5.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	88225	5.40	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.00%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	3591	0.191 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102120\
Data File : VU040900.D
Acq On : 22 Oct 2020 11:50
Operator : SY/MD
Sample : VU1021WBL02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK07

Quant Time: Oct 23 00:07:40 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 22 23:17:57 2020
Response via : Initial Calibration

