

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102218\
 Data File : VU027734.D
 Acq On : 22 Oct 2018 11:15
 Operator : MD/SY
 Sample : VU1022WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK62

Quant Time: Oct 23 08:13:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM102218WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 23 08:01:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	119827	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	108676	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	48110	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	44068	47.27	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.54%
7) Chloroethane-d5	1.68	69	34304	47.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.58%
11) 1,1-Dichloroethene-d2	2.27	63	61306	36.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.02%
21) 2-Butanone-d5	4.18	46	76037	95.70	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.70%
24) Chloroform-d	4.65	84	78140	46.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.24%
26) 1,2-Dichloroethane-d4	5.31	65	55841	48.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.22%
32) Benzene-d6	5.34	84	155291	50.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.36%
36) 1,2-Dichloropropane-d6	6.33	67	55501	50.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.46%
41) Toluene-d8	7.57	98	142827	50.03	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.06%
43) trans-1,3-Dichloropropene-	7.85	79	27223	50.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.46%
47) 2-Hexanone-d5	8.31	63	55433	98.03	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.03%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	68607	48.08	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.16%
64) 1,2-Dichlorobenzene-d4	11.86	152	49079	49.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102218\
Data File : VU027734.D
Acq On : 22 Oct 2018 11:15
Operator : MD/SY
Sample : VU1022WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK62

Quant Time: Oct 23 08:13:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM102218WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 23 08:01:28 2018
Response via : Initial Calibration

