

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102219\
 Data File : VU035328.D
 Acq On : 21 Oct 2019 18:03
 Operator : JC/SP
 Sample : VU1022WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK28

Quant Time: Oct 22 06:48:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 22 06:40:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	339545	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	321038	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	118431	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	124469	6.01	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	120.20%
7) Chloroethane-d5	1.93	69	110330	5.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	116.60%
11) 1,1-Dichloroethene-d2	2.59	63	168889	4.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.20%
20) 2-Butanone-d5	4.70	46	270666	45.49	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.98%
24) Chloroform-d	5.11	84	195715	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.40%
26) 1,2-Dichloroethane-d4	5.75	65	107411	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.40%
32) Benzene-d6	5.77	84	428541	5.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	116.80%
36) 1,2-Dichloropropane-d6	6.74	67	131293	5.63	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	112.60%
41) Toluene-d8	7.93	98	390189	5.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.20%
43) trans-1,3-Dichloropropene-	8.22	79	51400	4.54	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.80%
46) 2-Hexanone-d5	8.69	63	218833	43.86	ug/L	0.01
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.72%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	99086	4.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.40%
64) 1,2-Dichlorobenzene-d4	12.23	152	115186	5.34	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102219\
Data File : VU035328.D
Acq On : 21 Oct 2019 18:03
Operator : JC/SP
Sample : VU1022WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK28

Quant Time: Oct 22 06:48:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 22 06:40:31 2019
Response via : Initial Calibration

