

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U102219\
 Data File : VU035340.D
 Acq On : 21 Oct 2019 23:24
 Operator : JC/SP
 Sample : K5421-12
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AC3

Quant Time: Oct 22 07:46:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 22 06:40:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	282466	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	277545	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	99707	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	76598	4.45	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.00%
7) Chloroethane-d5	1.94	69	77438	4.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.40%
11) 1,1-Dichloroethene-d2	2.59	63	111347	3.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.00%
20) 2-Butanone-d5	4.71	46	248287	50.16	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.32%
24) Chloroform-d	5.12	84	159298	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.20%
26) 1,2-Dichloroethane-d4	5.75	65	90381	5.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.60%
32) Benzene-d6	5.77	84	321439	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.20%
36) 1,2-Dichloropropane-d6	6.74	67	107121	5.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.20%
41) Toluene-d8	7.94	98	283129	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.20%
43) trans-1,3-Dichloropropene-	8.22	79	40790	4.17	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.40%
46) 2-Hexanone-d5	8.69	63	194849	45.18	ug/L	0.01
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.36%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	91897	5.12	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.40%
64) 1,2-Dichlorobenzene-d4	12.22	152	92425	5.09	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.80%

Target Compounds

					Ovalue
13) Acetone	2.67	43	14492	4.209	ug/L 99
16) Methylene chloride	3.09	84	3918	0.179	ug/L 85
42) Toluene	8.01	91	18297	0.203	ug/L 98
53) m,p-Xylene	9.73	106	3275	0.085	ug/L 85
54) o-Xylene	10.13	106	2281	0.060	ug/L 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102219\
Data File : VU035340.D
Acq On : 21 Oct 2019 23:24
Operator : JC/SP
Sample : K5421-12
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AC3

Quant Time: Oct 22 07:46:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 22 06:40:31 2019
Response via : Initial Calibration

