

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U102219\
 Data File : VU035353.D
 Acq On : 22 Oct 2019 10:17
 Operator : JC/SP
 Sample : K5421-14
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Oct 24 13:34:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 23 06:48:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	263535	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	268392	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	102563	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	68653	4.27	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.40%
7) Chloroethane-d5	1.93	69	70592	4.81	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.20%
11) 1,1-Dichloroethene-d2	2.59	63	96536	3.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.00%
20) 2-Butanone-d5	4.72	46	263518	57.07	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.14%
24) Chloroform-d	5.11	84	152608	5.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.00%
26) 1,2-Dichloroethane-d4	5.75	65	86411	5.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.20%
32) Benzene-d6	5.77	84	285188	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.00%
36) 1,2-Dichloropropane-d6	6.74	67	101004	5.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.60%
41) Toluene-d8	7.93	98	241730	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.40%
43) trans-1,3-Dichloropropene-	8.22	79	39225	4.15	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.00%
46) 2-Hexanone-d5	8.69	63	200436	48.06	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.12%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	92579	5.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.60%
64) 1,2-Dichlorobenzene-d4	12.22	152	89785	4.81	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102219\
Data File : VU035353.D
Acq On : 22 Oct 2019 10:17
Operator : JC/SP
Sample : K5421-14
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Oct 24 13:34:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 23 06:48:58 2019
Response via : Initial Calibration

