

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102219\
 Data File : VU035366.D
 Acq On : 22 Oct 2019 16:14
 Operator : JC/SP
 Sample : VU1022WBL03
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK30

Quant Time: Oct 23 08:09:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 23 08:02:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	261234	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	265629	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	94322	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	72053	4.52	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.40%
7) Chloroethane-d5	1.93	69	71325	4.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.00%
11) 1,1-Dichloroethene-d2	2.59	63	96623	3.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.60%
20) 2-Butanone-d5	4.70	46	285423	62.35	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.70%
24) Chloroform-d	5.11	84	153267	5.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.40%
26) 1,2-Dichloroethane-d4	5.75	65	89333	5.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.80%
32) Benzene-d6	5.77	84	286437	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
36) 1,2-Dichloropropane-d6	6.74	67	101100	5.24	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.80%
41) Toluene-d8	7.93	98	239465	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.40%
43) trans-1,3-Dichloropropene-	8.21	79	38663	4.13	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.60%
46) 2-Hexanone-d5	8.68	63	223160	54.06	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.12%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	95488	5.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.00%
64) 1,2-Dichlorobenzene-d4	12.22	152	82918	4.83	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102219\
Data File : VU035366.D
Acq On : 22 Oct 2019 16:14
Operator : JC/SP
Sample : VU1022WBL03
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK30

Quant Time: Oct 23 08:09:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 23 08:02:08 2019
Response via : Initial Calibration

