


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU102220\
 Data File : VU040910.D
 Acq On : 22 Oct 2020 17:48
 Operator : SY/MD
 Sample : BFB
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\524U102220DW.M
 Title : METHOD 524.2 VOLATILES DRINKING WATER
 Last Update : Fri Oct 23 07:07:33 2020



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2875

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.7	34752	PASS
75	95	30	60	56.0	89530	PASS
95	95	100	100	100.0	159957	PASS
96	95	5	9	6.8	10826	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	71.3	113981	PASS
175	174	5	9	7.4	8401	PASS
176	174	95	101	98.6	112336	PASS
177	176	5	9	6.0	6735	PASS