

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102221\
 Data File : VU045356.D
 Acq On : 22 Oct 2021 11:37
 Operator : SY/MD
 Sample : VU1022MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1022MBL01

Quant Time: Oct 22 15:10:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	201038	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	370432	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	342576	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	144034	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	156741	46.837	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.680%
35) Dibromofluoromethane	5.292	113	117371	50.227	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.460%
50) Toluene-d8	7.903	98	467097	50.589	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.180%
62) 4-Bromofluorobenzene	10.636	95	160121	42.005	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102221\
Data File : VU045356.D
Acq On : 22 Oct 2021 11:37
Operator : SY/MD
Sample : VU1022MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1022MBL01

Quant Time: Oct 22 15:10:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

