

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102221\
 Data File : VU045357.D
 Acq On : 22 Oct 2021 12:01
 Operator : SY/MD
 Sample : VU1022WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1022WBL01

Quant Time: Oct 22 15:10:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	209919	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	393525	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	363703	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	154246	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	166154	47.549	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.100%
35) Dibromofluoromethane	5.292	113	123709	49.833	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.660%
50) Toluene-d8	7.899	98	484068	49.351	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.700%
62) 4-Bromofluorobenzene	10.639	95	175597	43.362	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102221\
Data File : VU045357.D
Acq On : 22 Oct 2021 12:01
Operator : SY/MD
Sample : VU1022WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1022WBL01

Quant Time: Oct 22 15:10:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

