

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102221\
 Data File : VU045361.D
 Acq On : 22 Oct 2021 13:49
 Operator : SY/MD
 Sample : M4288-01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-01-101921

Quant Time: Oct 22 15:11:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	225950	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	416336	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.427	117	393867	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.825	152	182822	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	173006	45.997	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.000%
35) Dibromofluoromethane	5.304	113	121449	46.242	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.480%
50) Toluene-d8	7.902	98	527404	50.823	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.640%
62) 4-Bromofluorobenzene	10.648	95	193469	45.158	ug/l	0.01
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.320%
Target Compounds						
68) m/p-Xylenes	9.703	106	12710	2.153	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102221\
Data File : VU045361.D
Acq On : 22 Oct 2021 13:49
Operator : SY/MD
Sample : M4288-01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-01-101921

Quant Time: Oct 22 15:11:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

