

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102221\
 Data File : VU045367.D
 Acq On : 22 Oct 2021 16:11
 Operator : SY/MD
 Sample : M4315-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MH-DD-D

Quant Time: Oct 23 01:12:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	215269	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	406578	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	376832	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	170708	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	170711	47.639	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.280%
35) Dibromofluoromethane	5.292	113	125094	48.773	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.540%
50) Toluene-d8	7.899	98	503543	49.688	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.380%
62) 4-Bromofluorobenzene	10.635	95	189349	45.257	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.520%
Target Compounds						
					Qvalue	
16) Acetone	2.639	43	14460	6.798	ug/l	99
18) Methyl Acetate	2.961	43	4883	0.785	ug/l	92
20) Methylene Chloride	3.047	84	16392	5.116	ug/l	96
43) Isopropyl Acetate	5.915	43	30180	3.566	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102221\
Data File : VU045367.D
Acq On : 22 Oct 2021 16:11
Operator : SY/MD
Sample : M4315-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MH-DD-D

Quant Time: Oct 23 01:12:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

