

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102318\
 Data File : VU027755.D
 Acq On : 23 Oct 2018 08:48
 Operator : MD/SY
 Sample : VU1023WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1023WBL01

Quant Time: Oct 24 05:26:24 2018
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U0102318DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Oct 23 09:12:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	5.75	96	14693	1.00	ug/l	0.00

System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.31	95	5160	0.93	ug/l	0.00
Spiked Amount 1.000			Recovery	=	93.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	4914	0.93	ug/l	0.00
Spiked Amount 1.000			Recovery	=	93.00%	

Target Compounds	Ovalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102318\
Data File : VU027755.D
Acq On : 23 Oct 2018 08:48
Operator : MD/SY
Sample : VU1023WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1023WBL01

Quant Time: Oct 24 05:26:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U0102318DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Oct 23 09:12:15 2018
Response via : Initial Calibration

