

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102318\
 Data File : VU027760.D
 Acq On : 23 Oct 2018 12:58
 Operator : MD/SY
 Sample : J5026-05DL 4X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 1720DL

Quant Time: Oct 24 05:54:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U0102318DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Oct 23 09:12:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	5.75	96	12784	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.31	95	4475	0.92	ug/l	0.00
Spiked Amount 1.000			Recovery	=	92.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	4266	0.93	ug/l	0.00
Spiked Amount 1.000			Recovery	=	93.00%	
Target Compounds						Ovalue
2) Dichlorodifluoromethane	1.21	85	41169	8.110	ug/l	100
4) Vinyl Chloride	1.40	62	36063	7.118	ug/l	98
5) Bromomethane	1.63	94	12651	4.747	ug/l	98
6) Chloroethane	1.70	64	9236	3.362	ug/l	95
7) Trichlorofluoromethane	1.89	101	35143	5.691	ug/l	99
9) 1,1-Dichloroethene	2.29	96	13228	4.639	ug/l	98
15) Methylene Chloride	2.70	84	10816	3.010	ug/l	97
16) trans-1,2-Dichloroethene	2.99	96	2449	0.772	ug/l	93
22) cis-1,2-Dichloroethene	4.22	96	15222	4.006	ug/l	89
25) Methyl tert-Butyl Ether	3.01	73	77626	8.550	ug/l	100
28) 1,1,1-Trichloroethane	4.91	97	4853	0.799	ug/l	99
29) 1,1-Dichloropropene	5.13	75	19848	3.744	ug/l	99
30) Carbon Tetrachloride	5.14	117	5012	0.924	ug/l	97
36) 1,2-Dichloroethane	5.41	62	19718	3.999	ug/l	98
37) Trichloroethene	6.19	130	9680	2.648	ug/l	94
38) 1,2-Dichloropropane	6.44	63	16505	3.967	ug/l	97
50) t-1,3-Dichloropropene	7.88	75	17858	3.490	ug/l	97
51) cis-1,3-Dichloropropene	7.27	75	12724	2.111	ug/l	99
52) 1,1,2-Trichloroethane	8.07	97	3216	1.259	ug/l	95
53) 1,3-Dichloropropane	8.25	76	16825	3.428	ug/l	99
58) Tetrachloroethene	8.23	164	10000	2.901	ug/l	96
60) 1,1,1,2-Tetrachloroethane	9.21	131	1773	0.518	ug/l	97
71) 1,2,3-Trichloropropane	10.49	75	12499	4.736	ug/l #	99
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102318\
Data File : VU027760.D
Acq On : 23 Oct 2018 12:58
Operator : MD/SY
Sample : J5026-05DL 4X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
1720DL

Quant Time: Oct 24 05:54:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U0102318DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Oct 23 09:12:15 2018
Response via : Initial Calibration

