

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102319\
 Data File : VU035432.D
 Acq On : 24 Oct 2019 02:40
 Operator : JC/SP
 Sample : VU1023WBL03
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK04

Quant Time: Oct 24 07:59:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 24 07:51:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	236544	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	243499	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	85234	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	85713	5.94	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	118.80%
7) Chloroethane-d5	1.93	69	78507	5.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.00%
11) 1,1-Dichloroethene-d2	2.59	63	112820	4.23	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.60%
20) 2-Butanone-d5	4.71	46	235117	56.72	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.44%
24) Chloroform-d	5.11	84	153137	5.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	116.20%
26) 1,2-Dichloroethane-d4	5.75	65	85351	5.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	116.80%
32) Benzene-d6	5.77	84	289916	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.20%
36) 1,2-Dichloropropane-d6	6.74	67	99755	5.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	112.80%
41) Toluene-d8	7.93	98	240379	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.20%
43) trans-1,3-Dichloropropene-	8.22	79	33489	3.90	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.00%
46) 2-Hexanone-d5	8.68	63	163456	43.20	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.40%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	85250	5.41	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.20%
64) 1,2-Dichlorobenzene-d4	12.22	152	79402	5.11	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102319\
Data File : VU035432.D
Acq On : 24 Oct 2019 02:40
Operator : JC/SP
Sample : VU1023WBL03
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK04

Quant Time: Oct 24 07:59:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 24 07:51:06 2019
Response via : Initial Calibration

