

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U102319\
 Data File : VU035443.D
 Acq On : 24 Oct 2019 07:37
 Operator : JC/SP
 Sample : K5509-06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H0G46

Quant Time: Oct 24 08:46:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 23 08:45:49 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	196873	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	204842	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.85	152	60201	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	49738	4.14	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.80%
7) Chloroethane-d5	1.93	69	53336	4.86	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.20%
11) 1,1-Dichloroethene-d2	2.60	63	72083	3.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.00%
20) 2-Butanone-d5	4.70	46	221786	64.29	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	128.58%
24) Chloroform-d	5.11	84	121113	5.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.40%
26) 1,2-Dichloroethane-d4	5.75	65	71566	5.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.80%
32) Benzene-d6	5.77	84	214951	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.80%
36) 1,2-Dichloropropane-d6	6.74	67	81242	5.46	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.20%
41) Toluene-d8	7.93	98	174249	3.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.80%
43) trans-1,3-Dichloropropene-	8.22	79	27301	3.78	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.60%
46) 2-Hexanone-d5	8.70	63	150740	47.35	ug/L	0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.70%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	75776	5.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	114.40%
64) 1,2-Dichlorobenzene-d4	12.23	152	56199	5.13	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102319\
Data File : VU035443.D
Acq On : 24 Oct 2019 07:37
Operator : JC/SP
Sample : K5509-06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H0G46

Quant Time: Oct 24 08:46:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR101719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 23 08:45:49 2019
Response via : Initial Calibration

