

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102320\
 Data File : VU040927.D
 Acq On : 23 Oct 2020 10:28
 Operator : SY/MD
 Sample : VU1023WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1023WBL01

Quant Time: Oct 24 06:38:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U102220DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Oct 23 07:07:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	6.12	96	68012	1.00	ug/l	0.00

System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.63	95	17686	0.71	ug/l	0.00
Spiked Amount 1.000			Recovery	=	71.00%	
68) 1,2-Dichlorobenzene-d4	12.19	152	18772	0.72	ug/l	0.00
Spiked Amount 1.000			Recovery	=	72.00%	

Target Compounds	Ovalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102320\
Data File : VU040927.D
Acq On : 23 Oct 2020 10:28
Operator : SY/MD
Sample : VU1023WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1023WBL01

Quant Time: Oct 24 06:38:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U102220DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Oct 23 07:07:33 2020
Response via : Initial Calibration

