

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102323\
 Data File : VU055849.D
 Acq On : 23 Oct 2023 12:38
 Operator : MD/SY
 Sample : VU1023MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK113

Quant Time: Oct 23 23:18:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM101023WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Oct 23 01:04:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	466246	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	466250	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	223721	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	160862	38.465	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	76.920%	
7) Chloroethane-d5	1.895	69	147557	43.164	ug/L	-0.01
Spiked Amount	50.000	Range 70 - 130	Recovery	=	86.320%	
11) 1,1-Dichloroethene-d2	2.560	63	240698	32.279	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	64.560%	
21) 2-Butanone-d5	4.615	46	247018	79.713	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	79.710%	
24) Chloroform-d	5.055	84	366598	45.726	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.460%	
26) 1,2-Dichloroethane-d4	5.695	65	236815	47.989	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	95.980%	
32) Benzene-d6	5.721	84	721361	48.355	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.720%	
36) 1,2-Dichloropropane-d6	6.685	67	233356	49.412	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	98.820%	
41) Toluene-d8	7.894	98	629852	46.758	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	93.520%	
43) trans-1,3-Dichloroprop...	8.177	79	100331	44.225	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	88.460%	
47) 2-Hexanone-d5	8.631	63	120378	64.837	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	64.840%	
56) 1,1,2,2-Tetrachloroeth...	10.753	84	360762	49.628	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	99.260%	
66) 1,2-Dichlorobenzene-d4	12.190	152	245766	55.357	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	110.720%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102323\
Data File : VU055849.D
Acq On : 23 Oct 2023 12:38
Operator : MD/SY
Sample : VU1023MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK113

Quant Time: Oct 23 23:18:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM101023WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Oct 23 01:04:15 2023
Response via : Initial Calibration

