

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU102418\
 Data File : VU027794.D
 Acq On : 24 Oct 2018 19:00
 Operator : MD/SY
 Sample : J5599-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H1A02

Quant Time: Oct 25 07:46:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102418WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 25 07:21:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	37466	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	36995	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	18008	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	13920	4.56	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.20%
7) Chloroethane-d5	1.68	69	10991	4.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.20%
11) 1,1-Dichloroethene-d2	2.28	63	20537	3.52	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.40%
20) 2-Butanone-d5	4.20	46	48312	52.03	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.06%
24) Chloroform-d	4.65	84	23482	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.20%
26) 1,2-Dichloroethane-d4	5.31	65	15970	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.40%
32) Benzene-d6	5.34	84	48467	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.20%
36) 1,2-Dichloropropane-d6	6.33	67	16801	4.80	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.00%
41) Toluene-d8	7.57	98	45429	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.60%
43) trans-1,3-Dichloropropene-	7.85	79	6662	4.69	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.80%
46) 2-Hexanone-d5	8.31	63	40288	51.41	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.82%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	12897	4.85	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.00%
64) 1,2-Dichlorobenzene-d4	11.86	152	17272	5.09	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU102418\
Data File : VU027794.D
Acq On : 24 Oct 2018 19:00
Operator : MD/SY
Sample : J5599-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H1A02

Quant Time: Oct 25 07:46:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102418WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 25 07:21:42 2018
Response via : Initial Calibration

