

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU102418\
 Data File : VU027797.D
 Acq On : 24 Oct 2018 20:12
 Operator : MD/SY
 Sample : J5599-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H1A05

Quant Time: Oct 25 08:08:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102418WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 25 07:21:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	39897	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	38334	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	19246	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	15976	4.91	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.20%
7) Chloroethane-d5	1.68	69	12619	5.13	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.60%
11) 1,1-Dichloroethene-d2	2.28	63	22266	3.58	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.60%
20) 2-Butanone-d5	4.19	46	50577	51.15	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.30%
24) Chloroform-d	4.65	84	26499	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
26) 1,2-Dichloroethane-d4	5.31	65	16346	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.60%
32) Benzene-d6	5.34	84	51082	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.00%
36) 1,2-Dichloropropane-d6	6.33	67	18485	5.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.00%
41) Toluene-d8	7.57	98	46849	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
43) trans-1,3-Dichloropropene-	7.85	79	6938	4.71	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.20%
46) 2-Hexanone-d5	8.31	63	41838	51.52	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.04%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	13726	4.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	18023	4.97	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.40%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU102418\
Data File : VU027797.D
Acq On : 24 Oct 2018 20:12
Operator : MD/SY
Sample : J5599-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H1A05

Quant Time: Oct 25 08:08:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102418WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 25 07:21:42 2018
Response via : Initial Calibration

