

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU102418\
 Data File : VU027800.D
 Acq On : 24 Oct 2018 21:24
 Operator : MD/SY
 Sample : J5599-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H1A08

Quant Time: Oct 25 08:14:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102418WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 25 07:21:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	37543	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	36184	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	18000	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	13200	4.31	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.20%
7) Chloroethane-d5	1.68	69	10949	4.73	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.60%
11) 1,1-Dichloroethene-d2	2.28	63	20814	3.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.20%
20) 2-Butanone-d5	4.19	46	48499	52.12	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.24%
24) Chloroform-d	4.65	84	24550	4.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.00%
26) 1,2-Dichloroethane-d4	5.31	65	16442	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.20%
32) Benzene-d6	5.34	84	49128	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.80%
36) 1,2-Dichloropropane-d6	6.33	67	17760	5.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.80%
41) Toluene-d8	7.57	98	45960	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.00%
43) trans-1,3-Dichloropropene-	7.85	79	6791	4.89	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.80%
46) 2-Hexanone-d5	8.31	63	40170	52.41	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.82%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	13318	5.12	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.40%
64) 1,2-Dichlorobenzene-d4	11.86	152	17352	5.12	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.40%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU102418\
Data File : VU027800.D
Acq On : 24 Oct 2018 21:24
Operator : MD/SY
Sample : J5599-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H1A08

Quant Time: Oct 25 08:14:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102418WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 25 07:21:42 2018
Response via : Initial Calibration

